

Advanced Cancer Institute

(Baba Farid University of Health Sciences, Faridkot)

B-15, Industrial Growth Centre,

Mansa Road Institute, Bathinda (Punjab) PIN 151001

Email: acdtrc@rediffmail.com Phone No: +91-164-2430418

No.ACI/BTI/2022/11354

Dated. 30.12.22.....

Incharge, IT cell
BFUHS, FDK

Subject: Quotation invited for Blood Mixer(Roller mixer) required at ACI, Bathinda


Sealed quotations are invited for Blood mixer(Roller Mixer) from different supplier which is required in this Institution with the given Terms & Conditions and specifications:-

S. No	Particulars	Qty.	Specifications
1.	Blood Mixer(Roller Mixer)	01	Roller size- 4 Nos.x230 mm Lg RPM Range-30 /36 RPM Power-230v,50 Hz/110v,60 Hz,AC Mixing Time: 15 to 20 minutes Weight: 2.0 Kg. Net Capacity: vaccutainer 6 Nos.(of – 4ml) Or Bulb of 5mlx15 Nos

Terms & Conditions:

1. The bidders should have to follow all the Terms & Conditions.
2. The material/equipment should be of Good Quality and of reputed brand/make.
3. Supply should be **F.O.R** Destination
4. Taxes/GST should be clearly mentioned and as per Govt. guidelines.
5. The quantity may increase or decrease according to the requirement.
6. Guarantee/Warranty should be quoted.
7. Payment will be released after satisfactory report from the concerned deptt./Stores.
8. Director, ACI, Bathinda reserves all rights to reject any Quotation/material without assigning any reason.
9. The material should be supplied within 15 days from the date of issue supply order otherwise 2% LD charges will be levied on the price
The bidder should submit the quotation in sealed envelope, super subscribing **“Quotation for Blood Mixer”** in favour of **The Director**, Advanced Cancer Institute, Bathinda

The Last Date for receipt of Quotation in the Director Office is 20.01.2023 by 3.00 PM


Director