

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

E-Tender Form

(E-Tender enquiry for Supply and Installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.)

Tender Notification No :	To be provided by the E-procurement portal of the Govt. of Punjab.
Tender Notification Date:	17/07/2021
Requirement	E-Tender notice for supply and Installation of OT Table, Anaesthesia Work Station and OT Lights
Cost of the tender document:-	Rs.2360/- (Non-refundable) to be deposited through Online Mode Only in favor of Registrar, Baba Farid University of Health Sciences, Faridkot.
Tender Processing Fee	To be charged by Govt. of Punjab as per its norms. (Non- refundable)
Earnest Money Deposit (EMD)	As mentioned in Scope of Supply. The Earnest Money Deposit must be submitted in the shape of Online Payment in favor of Registrar, Baba Farid University of Health Sciences, Faridkot on or before due date (Refundable to the Non-successful bidders, without any type of interest or other charges). In case of successful tenderer , it will be returned after receipt of the security amount as per tender terms and conditions.
Date of start of downloading of tender documents	28.07.2021 from the website of the Punjab Government i.e. https://eproc.punjab.gov.in
Website for downloading of the tender document:-	https://eproc.punjab.gov.in However, the details may also be obtained from the University website i.e. www.bfuhs.ac.in and college website www.ggsmch.org
Last date for downloading of the tender document:-	03.08.2021 up to 12.30 pm
Last date & time for uploading of the tender documents:-	03.08.2021 up to 1.30 pm (through online mode only)
Date, time and venue for opening of the Technical Bids	By the next day from the last date of submission of tenders (by 5:00 p.m.) on the e- procurement portal of the Govt. of Punjab. However the bidder or their authorized agents / representative may attend the tender opening process at the University Procurement & Facility Department, Baba Farid University of Health Sciences, Faridkot on the prescribed time and date.
Date, time and venue for opening of the Price Bids	The opening date of financial bids of the technically qualified bidders will be informed on the university web site, e- procurement portal of the Govt. of Punjab. However the bidder or their authorized agents / representative may attend the tender opening process at the University Procurement & Facility Department, Baba Farid University of Health Sciences, Faridkot on the prescribed time and date. The University authorities will have right to decide any other date and time for the opening of the Financial bids.
Who can be contacted for obtaining more information about the tender.	Principal, Guru Gobind Singh Medical College & Hospital, Sadiq Road, Faridkot. 01639-251111, 98773-65600, 94655-13138 E-mail: procurement@ggsmch.org ggsmc@punjab.gov.in , (on all working days from 9.00 a.m. to 5.00 p.m.)

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

NOTICE INVITING E-TENDER

E-Tenders are invited on or before **03.08.2021** from manufacturers or their authorized agents/distributors **for supply and Installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.** The tender document containing detailed terms & conditions may be downloaded from the E-procurement website of the Punjab Government i.e. <https://eproc.punjab.gov.in> and its detail may also be seen at the University website www.bfuhs.ac.in and college website www.ggsmch.org

TERMS AND CONDITIONS:-

1. The tender must be uploaded on or before the last date/ time of the submission of tender.
2. The Tender processing fee should be submitted through Net Banking/Credit card/Online mode only and as per Punjab Govt.
3. The **tenders will be opened online** by the next day from the last date of submission of tenders (by 5:00 p.m.) on the website i.e. <https://eproc.punjab.gov.in> at the, Baba Farid University of Health Sciences, Faridkot. The bidder(s) shall be at liberty to be present, in person or through their authorized representative(s) at the time of opening of the tender as specified in the Tender Notice. In case the authorized representatives are to be present, they must furnish the authority letter from the bidder (s), on whose behalf they are representing otherwise they will not be allowed to participate in the process of opening of tender.
4. **The Price bids of technically qualified bidders will be opened** on the website i.e. <https://eproc.punjab.gov.in> at, Baba Farid University of Health Sciences, Faridkot. In case of any change of date and time it will be notified to the technically qualified bidders through E-mail/telephone.
5. The Registrar/Principal reserves all rights to accept or reject any or all the tenders without assigning any reason.

Registrar

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

INSTRUCTIONS/ GUIDELINES TO THE TENDERERS

1. The bidder needs to register himself/ herself on <https://eproc.punjab.gov.in> the bidder is also required to obtain Class III digital signature certificates to complete this process.
2. Please download the Tender document from the website of e-procurement of the Govt. of Punjab <https://eproc.punjab.gov.in> Please fill all the relevant blanks on all the pages of the tender document sign along with a stamp/ seal all pages and then a scanned copy of the same may be uploaded on the website at the time of submission of the tender document.
3. **It should be clearly noted that this tender will be accepted though e-tender mode only.** The tenders submitted through offline mode will not be accepted under any circumstances.
4. **Tender Fee :** as per Govt. and should be deposited through online mode only
5. **Tender Processing Fee:** as per Punjab Govt. (non-refundable) may be deposited through online mode i.e. Net Banking/ Credit Card/ Debit Card only. The tender processing fee will not be accepted through any other mode.
6. **Refundable Earnest Money Deposit (EMD):** As mentioned in Scope of Supply to be deposited online mode only in favor of Registrar Baba Farid University of Health Sciences, Faridkot.
7. **Upload** signed copy of Technical Bid Compliance Statement (**Annexure-I**).
8. **Upload** an affidavit regarding Non-Black listing as per proforma given at **Annexure-II** duly attested by an Executive Magistrate or a Notary Public.
9. In case the Bidder is Authorized Supplier/Agency, the Authorization Certificate as per the Format given at **Annexure-‘III’** (duly filled in), **to be uploaded.**
10. In case the Bidder is Authorized Supplier/Agency, an undertaking/certificate issued by their Principal Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply of the Instruments then the Principal Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC of the Equipments/Instruments. (**Annexure – ‘IV’**), **to be uploaded.**
11. **Upload** details of Bank Account for refund of EMD (**Annexure – V**).
12. In addition to this, following **documents are to be uploaded** with Technical Bid:-
 - i) Details of registration as Company /Firm/ Establishment.
 - ii) **Certification as per technical specifications.**
 - iii) Copy of Certificate of Registration for service Tax/TIN/TAN/PAN/VAT.
 - iv) A certificate from C.A. regarding Annual Turnover with Balance Sheet for the last 3 (three) financial years i.e. 2017-18, 2018-19 & 2019-20.
 - v) Copy of the IT Returns for three financial years i.e. 2017-18, 2018-19 & 2019-20.
15. Price should be quoted and **uploaded** only in Excel Sheet proforma at **Annexure-‘VI’**.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

SCOPE OF SUPPLY

Sr.	Name of Equipment	Approx. Qty. Reqd.	Earnest Money to be Submitted (in Rs.)
1.	OT Table	04	1,50,000.00
2.	OT Lights	05	2,50,000.00
3.	Anaesthesia Work Station With Monitors	05	4,50,000.00

Quantity may increase / decrease as per requirement.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

TECHNICAL SPECIFICATIONS

OT Light

1. One Major and One Satellite Light Head of Different dimensions
2. There should be wireless HD camera with recorder and Monitor Arm for Mounting 28 - 31 inch HD screen (Fully HD).
3. Monitor Reflector based LED Technology for better shadow resolution
4. Arrangement of LED in such a way that Shadow Free / Deep cavity illumination is achieved.
5. Adjustable color temp Special design to maximize the field of illumination and optimized illumination depth Should have good laminar flow properties
6. Should have Aluminium Housing for better heat management
7. ESG safety glass for simple and fast disinfection process
8. 360 deg rotation of domes/lighthoods / arms for unlimited positioning of light heads
9. MIS Lightning Feature on domes
10. Multifunctional handle Should have control panel on the cardanic arms
11. In case of LED failure , each individual LED module should be replacable
12. Central illuminationb 160,000 lux or more
13. Life time of light source (h) \geq 60,000
14. Light Field Diameter (mm) 180-300
15. Colour Temp(K): 3500-5500 K
16. Preferable Chargeable Battery Backup atleast 1 hr
17. Manufacturer warranty 5years including LED lamps , electronic components and mechanical spares.
18. Depth of Illumination 120-140 cm or more.
19. Colour rendering index 93-98
20. Height Adjustment – more than one meter
21. Certification ISO/FDA-US/European Std.
22. CMC -5years after expiry of warranty

C –Arm compatible OT table

The table should be remote controlled electrically operated motorised OT table with battery backup system. The table should have an additional / auxiliary control

The Table should have at least four sections and table top should be radio lucent to permit X-RAY penetration and fluoroscopy.

The table should be designed for Orthopedic, Cardiology, Urology, Endoscopic Retrograde Cholangiopancreatogram (ERCP) and also equipped with a translucent top for C- Arm use.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Table should have motorized surgical positions like Height Adjustment, raising, lowering, trendelenburg, reverse trendelenburg, lateral tilt, flex, re- flex, Longitudinal Slide (300 mm), One touch Zero positioning, Electrical Floor Locking & chair position etc.

Seamless moulded mattress should be radiolucent material and suitable for fluoroscopy.

Table should have zero leveling facility.

In built Kidney Bridge supplied complete with standard accessories and mattress/ flex or extension position without Kidney Bridge.

The entire table should be made of stainless steel grade SS 304 and other high quality materials.

Table provided with SS- 304 Blocks and comes with other accessories including arm board, leg rest, mattress, shoulder support & one handle.

Radiolucent top allowing clear image on C-arm monitor.

Heavy frame provided for fixing of gears on the sides of table which do not interfere with the movement of C-arm.

Measurements (all dimensions are approximated to +/- 10% variations)

Top dimension	L 2140 X W 553 mm
Height adjustment	575mm X 1012 mm
Trendelenburg / Reverse	30 ⁰ / 25 ⁰
Lateral tilt	20 ⁰ / 20 ⁰
Kidney elevator	120 mm

Back Rest(up / down)	80 ⁰ / 25 ⁰
Leg Rest(up / down)	15 ⁰ / 90 ⁰
Head Rest(up / down)	20 ⁰ /60 ⁰

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

ORTHOAPEDIC AND SPINE ACCESSARIES FOR OT TABLE

SPINE FRAME / MATTRESSS

BOLSTER / PILLOWS 4 IN NO- PADDED/ MEMORY FOAM / EASY TO CLEAN

HORSE SHOE HEAD REST-Padded, contoured headrest

Versatile positioning in height, extension, and tilt angulations

Versatile for supine, prone, and fowler positioning

Durable, stainless steel construction

PRONE HEAD REST-Excellent support and protection of the patient's face.

Dual, side access for respiratory tubes

Anatomically specific contouring for prone positioning

Lightweight, ultra-soft,

CONNECTING BRACKET

STANDERED HOLDER FOR SKULL CLAMP- for cervical spine surgery

PAD PSOTIONERS- Had ring , heel pads , arm pads, table pads(aqua gel)

ARM SUPPPORT

D PILLOW WITH FIXING STRAPS FOR TKR SURGERY

SIDE SUPPORTS- Fully adjustable for use during total hip and other procedures requiring optimal lateral position stability

CLAMPS AND SOCKETS

ADULT SIDE ARM TABLE

LEG SUPPORT-Simple, versatile design

Ball joint with locking mechanism for flexibility and security

Contoured support with integral padding for comfort

Single application for the unaffected leg or can be used as a pair

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

MULTITASK ARM BOARD-Surgical arm board for lateral, chair, prone positions
Lockable ball socket permits a wide range of positioning flexibility

SURGICAL ORTHOPAEDIC TRACTION UNIT (HIP SPICA ASSEMBLY) WITH
PADDED PERINEAL POST

TIBIAL SUPPORT FOR NAILING

ORTHOPAEDIIC EXTENSSION CART

For neurosurgery table

Radiolucent May field/DORO skull clamp with modular fixation system- swivel adapter and long transitional member-one complete set including table attachment with storage container/s. Skull pins (reusable, autoclavable) adult 6, Pediatric 6
Radiolucent disposable pins adults-12, pediatric- 12 Accessories for park bench position-one set Accessories for sitting position- one set.

14. Standard, Safety and Training

- a. Should be US – FDA/European CE (from European union notified body having 4 digit identification number).
 - b. Manufacturer should be ISO certified for quality standards.
 - c. Should have local service facility. The service provider should have the necessary equipment.
15. Should be able to carry patient weight at least 250 kg.

Warranty – 5 years; CMC – 5 years.

16. Documentation

- a. User / Technical / Maintenance manuals to be supplied in English.
- b. Certificate of calibration and inspection. List of equipment available for providing calibration and routine Preventive Maintenance support as per manufacturer documentation in service / Technical manual.
- c. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Specification for Anaesthesia Workstation

1. General Requirement

- Compact and modular, three gas Anaesthesia workstation with an integrated ventilator for adult to infants and integrated airway monitor for airway pressures and volume. It should be constructed with welded tubular/epoxy powder painted steel or ABS medical grade plastic.
- The machine should be suitable for low and minimal flow anaesthesia application with compliance compensation of breathing circuit fresh gas flow compensation/ decoupling.
- The machine should have 2/3 drawers (preferably lockable)
- The anaesthesia machine, inbuilt ventilator and vaporizer should be manufactured by the single company
- The system should have upto 1 Hrs. battery backup
- System should conform to (US FDA / 4 digit European CE) and (BIS/ISO) and EN 60601-2-13 (Requirement for safety and essential performance of anaesthesia system)
- Top shelf and table top to keep drugs and equipment
- Should have large castor wheel with foot brake

2. Gas delivery system

- Should have pin index yokes for Oxygen & Nitrous Oxide besides separate connection for Central gas supply for Oxygen, Nitrous Oxide and Air.
- The machine should have pressure gauges for cylinders & central supply lines mounted on front of Anaesthesia machine for better visibility. The gas connections should be noninterchangeable.
- The system should be suitable to use at minimal flow upto 700ml fresh gas setting.
- Automatic cutoff of N₂O by Oxygen pressure failure.
- Hypoxic guard for linear regulation of minimum oxygen concentration at 23 - 25 % volume
- Audible visual oxygen failure alarm.
- Emergency Oxygen flush at 30 – 70 L/min bypassing the vaporizer.
- Should have analogue display of cylinders and pipeline gas pressures, should have provision for attaching one cylinder for nitrous oxide and one for oxygen cylinder.
- Auxillary oxygen port should be available

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

3. Flow meter

- Dual Cascade type flow meter tubes for Oxygen & N₂O. Range 20 ml / min to 10 Lit/min.
- Calibrated in multiple scales. Minimum one tube for air 100 ml to 14 L/ min.

4. Vaporizer

- Machine should have possibility to mount two quick mount type vaporizer for easy interchangeability, and safety with interlock facility.
- Should be provided with a Temperature / pressure compensated and flow independent
- Should be capable of connecting Vaporiser. for Isoflurane / Sevoflurane / Desflurane , Optional Price to be quoted for Halothene Vaporiser .
- Vaporizer should have extended delivery range from 0 to 8 Vol. %
- The vaporiser should be calibration free.

5. Breathing System

- Should have fresh gas de-coupled semi closed circle absorber system.
- Should have adjustable pressure relief valve from 0 to 75 mbar.
- The system should have leak and compliance test (including patient hoses upto the Y piece).
- Should have compact breathing system with approx 1.7 Ltr. Volume capacity.
- Should have an external fresh gas outlet for connecting Magill or Bain's circuit
- The device should have port for anaesthesia gas scavenging system.
- Should have fully autoclavable CO₂ absorber, canister

6. Anaesthesia Ventilator

- The system should have inbuilt ventilator with electronically controlled and Electronic/ pneumatic driven. Should not require changing of bellows for adult & infants.
- Should have minimum color screen size of 10" or more.
- Modes: Manual/Spont, Volume controlled, Pressure controlled, SIMP with PS
- Tidal Volume : 20 ~ 1400 ml
- PEEP : 0 ~ 20 mbar

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

- Breathing Frequency : 4 to 60 BPM
- I :E Ratio : 4:1 to 1:4
- Inspiratory pause: 0-50% of Ti.

7. Airway monitoring

- Integrated monitor for electronic monitoring and display of following parameters :
- Expiratory Tidal Volume
- Expiratory Minute volume
- PEEP, Peak & Mean and Plateau airway pressure
- Frequency
- Graphic representation for Lung mechanics

8. Alarm limits & alarms

- Adjustable high / low limits with audio and visual alarms for the following:
- Minute volume,
- Airway pressure (incl stenosis and disconnect),
- Insp oxygen concentration,
- Audio power supply fail alarm,
- Fail to cycle warning.

9. Patient Monitor

- The equipment should be advanced high end Patient monitoring system having integrated basic measurements & features with capability to provide additional clinical measurements by attaching single parameter modules
- The monitors should have bright, highly visible, minimum 12 XGA TFT Multicolor display touch Screen with rotator knob with minimum 6 waveforms
- The monitors should have the capability to provide ECG, Resp, NIBP, SpO₂, PVI, EtCO₂, 2 IBP/Temp, gas monitoring, BIS display.
- All monitors should have the capability to attach additional modules for the measurement of continuous 12 lead diagnostic ECG, Cardiac output both by thermo-dilution method as well as Continuous Cardiac Output , EEG, NMT, SvO₂.
- Upgradable to NICCO/PICCO module, web server

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Scope of supply

- 3 gas Anaesthesia machine with Patient Monitor
- Trolley with 2/3 drawers
- Pin Index yokes for O₂ & N₂O
- Pipeline connections for all three gases
- Anaesthesia ventilator
- Central gas supply hoses (Color coded) – 1 each
- Anaesthetic reusable silicon mask size – Adult & child- 5 set each (Size ; 00,0,1,2,3,4,5)
- Vaporisers for Isoflurane, Sevoflurane and Desflurane – 1 each
- 5 lead ECG Cable – 2 nos
- ECG 5 leads: 5 nos
- ECG disposable electrodes; 500
- Bains circuit (disposable): 10
- Jackson Rees circuit : 5
- Resusable closed circuit : 2 sets (adult and pediatric)
- Disposable breathing circuits: 10
- Rebreathing bags : 5 no (2 litre), 2 no (1 litre) and 3 no (500 ml)
- SpO₂ finger sensor with cables –4 no Adult and 4 no Paediatric (2 neonatal) per monitor
- Temperature probe: 2 adult (1 core, 1 skin) and 2 pediatric (1 core , 1 skin)
- Temperature cables –: 4 no
- NIBP Hose – 2 no
- NIBP cuff Adult & Paediatric – neonatal(3), pediatric(3), small (5), medium(5), large(5) and extra large(5)
- Should be supplied with intermediate IBP cable– 4 nos
- Disposable transducers – 30 nos
- Disposable gas sampling line – 50 nos
- Disposable water traps – 30 nos
- BIS module : 1 nos
- BIS cable : 2 nos
- BIS sensors- 50 nos

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

OTHERS:

1. Original Manufacturer or their subsidiary or authorized dealer who is quoting should be present in India having selling experience of 5 years .
2. In no case the instrument should remain in non-working condition for more than 7 days, beyond which a penalty as the per the rules of the institute will be levied
3. The vendor should have a good service and application back up along with instruments to provide an effective troubleshooting and support. (response time < 24 hours)
4. All technical bids comparative statement to the tender specifications must be enclosed along with reference no., paragraph no. from original catalogue of the equipment.
5. Demonstration and training should be provided at own expenditure.
6. All accessories like UPS / Stabilizer or any other which are required for machine should be supplied by the L1 bidder.
7. Compliance report to be submitted in a tabulated and point-wise manner, clearly mentioning the page/ para number with authenticated catalogue / manual. Points not covered in the brochure must be specifically addressed in a separate certificate.
8.
 - ❖ Certificate (s) regarding standard in quality must be uploaded.
 - ❖ Expected delivery & Installation time: Within 45 days after issue of supply order.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

TERMS AND CONDITIONS

FOR SUPPLY AND INSTALLATION of OT TABLE, ANAESTHESIA WORK STATION & OT LIGHTS AT GGS MEDICAL COLLEGE & HOSPITAL, FARIDKOT.

ELIGIBILITY

- The sole manufacturers of equipments or their authorized agents/distributors may quote their rates.
 - In case of Authorized Supplier/Agency/Distributor, the Authorization Certificate as per the Format given at **Annexure-‘III’** should be uploaded.
 - In case the Tenderer is authorized dealer/supplier an undertaking/certificate issued by their Principle Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply then the Principle Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC. (**Annexure – ‘IV’**).
1. This institution reserves the right to reject tenders without assigning any reason and increase or decrease the quantity of the articles tendered.
 2. *If the supply and installation* is not made within the stipulated period then late delivery charges @**2%** will be imposed on the total amount of Supply Order up to delay of **30 days** and thereafter @ **4%** for another **30 days** after which Supply Order will be deemed cancelled & security/earnest money will be forfeited and company will be black-listed for future.
 3. Payment Terms: 80% Payment will be released after satisfactory Installation of the Equipment and balance 20% will be made after 60 days of the Installation and satisfactory working of the equipment.
 4. In-complete or conditional offers incorporating price variation will not be entertained.
 5. The firm should have been in existence for at- least **three years** and it should have turnover of **Rs.2,00,00,000/- per year**.
 6. The successful bidder shall deposit performance security @ 10% of the basic cost of the equipment/s in the shape of **Demand Draft only** and will be returned after receipt of CAMC security.
 7. The successful bidder shall deposit CAMC security @ 10% of the total basic cost of the CAMC (05 years) in the shape of **Demand Draft only** and will be returned after 66 months.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Annexure-I

TECHNICAL BID COMPLIANCE STATEMENT

Name and Address of the applicant / firm _____

Specify whether Manufacturer/Dealer/Distributor: _____

Sr. No.	Particulars	Remarks
1.	Tender Fee of Rs.2360/- through Online Mode only on or before due date.	Yes/No
2.	Tender Processing fee charged by Punjab Govt. as per their norms. (Non-refundable).	Yes/No
3.	Earnest Money in favor of Registrar Baba Farid University of Health Sciences, Faridkot through online mode only.	Yes/No
4.	Technical Bid Compliance Proforma uploaded (Annexure-I).	Yes/No
5.	Whether an affidavit regarding Non-Black listing as per proforma given at Annexure-II duly attested by an Executive Magistrate or a Notary Public uploaded.	Yes/No
6.	In case the bidder is Authorized Supplier/Agency, the Authorization Certificate as per the Format given at Annexure-‘III’ uploaded.	Yes/No
7.	In case the Tenderer is Authorized Supplier/Agency, an undertaking/certificate issued by their Principle Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply then the Principle Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC (Annexure – ‘IV’) uploaded.	Yes/No
8.	Details of Bank Account for refund of EMD (Annexure – V) uploaded.	Yes/No
9.	Price Bid in the prescribed format in Excel Sheet (Annex – VI) uploaded.	Yes/No
10.	Copy of Certificate of Registration for service Tax/TIN/TAN/PAN uploaded.	Yes/No
11.	A certificate from C.A. regarding Annual Turnover with Balance Sheet for the last 3 (three) financial years i.e. 2017-18, 2018-19 and 2019-20 uploaded.	Yes/No
12.	Copy of the IT Returns for three financial i.e. 2017-18, 2018-19 and 2019-20 uploaded	Yes/No
13.	Certificate regarding standard in quality as per specifications	Yes/No
14.	E-mail ID	

Signature & seal of bidder

Place:

Date :

Note: Please upload Catalogue/Brochure/Pamphlets with complete specifications of quoted model only.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Annexure-II

(To be furnished on non-judicial stamp paper worth Rs.100/- duly attested by Executive Magistrate or Notary Public).

AFFIDAVIT

I/We _____
partner/sole proprietor (Strike out which is not applicable) of (Name & Address of Firm)
_____ do hereby declare and solemnly affirm:-

- a) That the individual/firm/ companies are **not debarred or black- listed** by any department of Union/ State Government or any autonomous institute.
- b) That no partner or shareholder, directly or indirectly connected with the applicant who has been debarred or blacklisted by any department of Union Govt./State Govt./Autonomous Institute.
- c) And that the terms and conditions for supply and Installation of Equipments at GGSMCH, Faridkot, are acceptable to me/us. I/We shall abide by them in letter and spirit.

Date:

Place:

DEPONENT

VERIFICATION

I/We do hereby solemnly declare and affirm that the above declarations are true and correct to the best of my/our knowledge and beliefs. No part of it is false and nothing has been concealed therein.

Date:

Place:

DEPONENT

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Annexure- III

MANUFACTURER'S/PRINCIPAL AUTHORIZATION FORM

TO

The Registrar
Baba Farid University of Health Sciences,
Faridkot -151203

Ref. No.....

Dated:

Sub: Authorization Certificate in favour of M/s..... for supply of (Name of equipment)

We, M/s....., who are established and reputable manufacturers of(name of equipment) having factory(ies) at and, hereby authorize M/s.....(name and address) to bid, negotiate and conclude the Tender formalities with you against Tender No..... for the above equipment(s) manufactured by us.

No company or firm or individual other than M/s..... are authorized to bid, negotiate and conclude the tender formalities in regard to this business against this specific tender.

We, hereby extend our full guarantee and warranty as per the conditions of tender for the goods offered for supply against this tender by the above firm.

Yours faithfully,

(Name)

For and on behalf of M/s.....
(name of manufacturer/Principal)

Note: This letter should be signed by a person competent and having authority to sign on behalf of manufacturer, and should be on manufacturer Letter Head and same will be uploaded with Technical Bid.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Annexure - IV

UNDERTAKING BY MANUFACTURER/PRINCIPAL SUPPLIER

TO

The Registrar
Baba Farid University of Health Sciences,
Faridkot -151203

Ref. No.....

Dated:

Sub: Undertaking for after sales service

We, M/s....., who are established and reputable manufacturers of(name of equipment) have authorized M/s.....(name and address) to bid, negotiate and conclude the Tender formalities with you against Tender No..... for the above equipment(s).

Further, we undertake that in case dealership/distributorship is withdrawn after supply of equipment then we shall be responsible for after sales service till the date of guarantee/warranty of the equipment and afterwards for a period of 10 years.

Yours faithfully,

(Name)

For and on behalf of M/s _____
(name of manufacturer/Principal)

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

Annexure- V

Details of Bank Account of the firm who has deposited EMD

Name of the firm: _____

Sr. No.	Particulars	Detail
1.	Account No.	
2.	Name of Bank	
3.	Branch Name	
4.	IFSC Code of Bank	
5.	Name of Operator	

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR supply & installation of OT Table, Anaesthesia Work Station and OT Lights at GGS Medical College & Hospital, Faridkot.

ANNEXURE - VI

PRICE BID

TO BE UPLOADED in Printed/Computerized format in Excel Sheet Attached for all Equipments.