

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine required at GGS Medical College & Hospital, Faridkot.

E-Tender Form

(E-Tender enquiry for Supply and Installation of X-ray Units of 100mA at GGS Medical College & Hospital, Faridkot.)

Tender Notification No :	To be provided by the E-procurement portal of the Govt. of Punjab.
Tender Notification Date:	02/07/2020
Requirement	E-Tender notice for supply and Installation of 100 mA X-ray Unit
Cost of the tender document:-	Rs.2000/- (Non-refundable) to be deposited through Online Mode Only in favor of Registrar, Baba Farid University of Health Sciences, Faridkot.
Tender Processing Fee	To be charged by Govt. of Punjab as per its norms. (Non- refundable)
Earnest Money Deposit (EMD)	Rs.15000/- (Rs. Fifteen thousand only). The Earnest Money Deposit must be submitted in the shape of Online Payment in favor of Registrar, Baba Farid University of Health Sciences, faridkot on or before due date (Refundable to the Non-successful bidders, without any type of interest or other charges).
Date of start of downloading of tender documents	02.07.2020 from the website of the Punjab Government i.e. https://eproc.punjab.gov.in
Website for downloading of the tender document:-	https://eproc.punjab.gov.in However, the details may also be obtained from the University website i.e. www.bfuhs.ac.in and college website www.ggsmch.org
Last date for downloading of the tender document:-	22.07.2020 up to 12.30 pm
Last date & time for uploading of the tender documents:-	22.07.2020 up to 1.30 pm (through online mode only)
Date, time and venue for opening of the Technical Bids	By the next day from the last date of submission of tenders (by 5:00 p.m.) on the e- procurement portal of the Govt. of Punjab in , Baba Farid University of Health Sciences, Faridkot
Date, time and venue for opening of the Price Bids	29.07.2020 on the e- procurement portal of the Govt. of Punjab in, Baba Farid University of Health Sciences, Faridkot
Who can be contacted for obtaining more information about the tender.	Principal, Guru Gobind Singh Medical College & Hospital, Sadiq Road, Faridkot. 01639-251111, 98773-65600, 79867-19101 E-mail: procurement@ggsmch.org (on all working days from 9.00 a.m. to 5.00 p.m.)

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine required at GGS Medical College & Hospital, Faridkot.

NOTICE INVITING E-TENDER

E-Tenders are invited on or before **22.07.2020** from manufacturers or their authorized agents/distributors **for supply and Installation of Portable HF 100mA X-ray Machine required at GGS Medical College & Hospital, Faridkot.** The tender document containing detailed terms & conditions may be downloaded from the E-procurement website of the Punjab Government i.e. <https://eproc.punjab.gov.in> and its detail may also be seen at the University website www.bfuhs.ac.in .

TERMS AND CONDITIONS:-

1. The tender must be uploaded on or before the last date/ time of the submission of tender.
2. The Tender processing fee should be submitted through Net Banking/Credit card/Online mode only and as per Punjab Govt.
3. The **tenders will be opened online** by the next day from the last date of submission of tenders (by 5:00 p.m.) on the website i.e. <https://eproc.punjab.gov.in> at the, Baba Farid University of Health Sciences, Faridkot. The bidder(s) shall be at liberty to be present, in person or through their authorized representative(s) at the time of opening of the tender as specified in the Tender Notice. In case the authorized representatives are to be present, they must furnish the authority letter from the bidder (s), on whose behalf they are representing otherwise they will not be allowed to participate in the process of opening of tender.
4. **The Price bids of technically qualified bidders will be opened** on **29.07.2020** at 2.30 p.m. on the website i.e. <https://eproc.punjab.gov.in> at, Baba Farid University of Health Sciences, Faridkot. In case of any change of date and time it will be notified to the technically qualified bidders through E-mail/telephone.
5. The Registrar/Principal reserves all rights to accept or reject any or all the tenders without assigning any reason.

Registrar

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine required at GGS Medical College & Hospital, Faridkot.

INSTRUCTIONS/ GUIDELINES TO THE TENDERERS

1. The bidder needs to register himself/ herself on <https://eproc.punjab.gov.in> the bidder is also required to obtain Class III digital signature certificates to complete this process.
2. Please download the Tender document from the website of e-procurement of the Govt. of Punjab <https://eproc.punjab.gov.in> Please fill all the relevant blanks on all the pages of the tender document sign along with a stamp/ seal all pages and then a scanned copy of the same may be uploaded on the website at the time of submission of the tender document.
3. **It should be clearly noted that this tender will be accepted though e-tender mode only.** The tenders submitted through offline mode will not be accepted under any circumstances.
4. **Tender Fee :** as per Govt. and should be deposited through online mode only
5. **Tender Processing Fee:** as per Punjab Govt. (non-refundable) may be deposited through online mode i.e. Net Banking/ Credit Card/ Debit Card only. The tender processing fee will not be accepted through any other mode.
6. **Refundable Earnest Money Deposit (EMD):** Rs.15,000/- to be deposited online mode only in favor of Registrar Baba Farid University of Health Sciences, Faridkot.
7. **Upload** signed copy of Technical Bid Compliance Statement (**Annexure-I**).
8. **Upload** an affidavit regarding Non-Black listing as per proforma given at **Annexure-II** duly attested by an Executive Magistrate or a Notary Public.
9. In case the Bidder is Authorized Supplier/Agency, the Authorization Certificate as per the Format given at **Annexure-‘III’** (duly filled in), **to be uploaded.**
10. In case the Bidder is Authorized Supplier/Agency, an undertaking/certificate issued by their Principal Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply of the Instruments then the Principal Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC of the Equipments/Instruments. (**Annexure – ‘IV’**), **to be uploaded.**
11. **Upload** details of Bank Account for refund of EMD (**Annexure – V**).
12. In addition to this, following **documents are to be uploaded** with Technical Bid:-
 - i) Details of registration as Company /Firm/ Establishment.
 - ii) **The system must have valid AERB Type approval, AERB certificate must enclosed with Bid.**
 - iii) Copy of Certificate of Registration for service Tax/TIN/TAN/PAN/VAT.
 - iv) A certificate from C.A. regarding Annual Turnover with Balance Sheet for the last 3 (three) financial years i.e. 2017-2018 , 2018-19 & 2019-20.
 - v) Copy of the IT Returns for three financial years i.e. 2017-2018 , 2018-19 & 2019-20.
13. Price should be quoted and **uploaded** only in Excel Sheet proforma at **Annexure-‘VI’**.

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.***

LIST OF EQUIPMENTS

Sr.	Name of Equipment	Approx. Qty. Reqd.	Earnest Money to be Submitted (in Rs.)
1.	Portable HF 100mA X-ray Machine	As per requirement	15000.00

Bidder must mention on firm's letter head:

- **Guarantee / Warranty**
- **Make & Model**
- **Projected delivery and installation time.**

TECHNICAL SPECIFICATIONS

Portable HF 100mA X-RAY MACHINE:

1. Compact, lightweight, easily transportable mobile High Frequency X-Ray unit for bedside x-Ray rays, trauma, Intensive care units, Operation theatres and Radiology Department.
2. The unit should be fully counter balanced and can be positioned to suit different bed heights. The unit should have facility of vertical swing and horizontal rotation of the tube head to ensure X- Ray of any anatomy even with in limited space.
3. The unit must have an effective braking system for parking and transport.
4. The exposure release switch should be detachable with a cord of sufficient length (at least 5 m)
5. **The Generator:**
 - a. Microprocessor controlled high frequency/inverter type of high frequency (80 KHz or more) for constant output.
 - b. It should have power rating of at least 4kW or more
 - c. It should have a digital display of mAs and kV.
 - d. KV range:40 kv to 100kV or more
 - e. mA range:10 mA to 100mA or more
 - f. KV selection: 40 kV to 100 kv, selectable in 1 kV steps
 - g. mAS selection: 0.1 to 250 mAS
 - h. It should have over loading protection.
 - i. It should have APR feature
6. **X-Ray Tube and Collimator:**
 - a. Stationary / Rotating anode having focal spot size less than 2mm
 - b. Output of tube should match with that of generator.
 - c. Light Beam diaphragm/ Double layer Collimator with auto cut off switch. The light intensity shall be at least 160 lux at1mtr distance from focal spot.
Heat storage capacity should be 200 KHU or more.

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.***

7. The unit should operate on single phase power supply and should have plug in facility to any standard wall outlet with automatic adaptation to line voltage 200 to 240volts, 15Amp plug.
8. The Leakage radiation level at 1 meter from the focus should be less than 70 mR. Products having minimal leakage radiation level will be preferred. (Please attach relevant test report)
9. The Systems should be fully safe with respect to
 - a. Over current
 - b. Overvoltage
 - c. Maximum loading of tube.
10. Power input to be 220-240VAC, 50Hz fitted with Indian plug.
11. The product should be CE or USFDA approved. Should be an AERB approved product. Two each 0.5mm lead equivalent thyroid shields and lead goggles to be supplied.
12. Vendor must provide complete machine cover for safety from dust while parked in the ward. Wheels of the machine must be of good quality material.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine required at GGS Medical College & Hospital, Faridkot.

TERMS AND CONDITIONS

FOR SUPPLY AND INSTALLATION of Portable HF 100mA X-ray Machine REQUIRED AT GGS MEDICAL COLLEGE & HOSPITAL, FARIDKOT.

ELIGIBILITY

- The sole manufacturers of equipments or their authorized agents/distributors may quote their rates.
 - In case of Authorized Supplier/Agency/Distributor, the Authorization Certificate as per the Format given at **Annexure-‘III’** should be uploaded.
 - In case the Tenderer is authorized dealer/supplier an undertaking/certificate issued by their Principle Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply then the Principle Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC. (**Annexure – ‘IV’**).
1. This institution reserves the right to reject tenders without assigning any reason and increase or decrease the quantity of the articles tendered.
 2. If the supply is not made within the stipulated period then late delivery charges @**2%** will be imposed on the total amount of Supply Order up to delay of **30 days** and thereafter @ **4%** for another **30 days** after which Supply Order will be deemed cancelled & security/earnest money will be forfeited and company will be black-listed for future.
 3. Payment Terms: 80% Payment will be released after satisfactory Installation and working of the Equipment and balance 20% will be made after 60 days of the Installation and satisfactory working of the equipment.
 4. In-complete or conditional offers incorporating price variation will not be entertained.
 5. The firm should have been in existence for at- least **three years** and it should have turn of **Rs.1,00,00,000/- per year**.
 6. The successful bidder shall deposit bank guarantee of 10% of the total value of the equipment/s in the shape of Demand Draft only which will be returned after the expiry of the guarantee.

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.***

Annexure-I

TECHNICAL BID COMPLIANCE STATEMENT

Name and Address of the applicant / firm _____

Specify whether Manufacturer/Dealer/Distributor: _____

Sr. No.	Particulars	Remarks
1.	Tender Fee of Rs.2000/- through Online Mode only on or before due date.	Yes/No
2.	Tender Processing fee charged by Punjab Govt. as per their norms. (Non-refundable).	Yes/No
3.	Earnest Money in favor of Registrar Baba Farid University of Health Sciences, Faridkot through online mode only.	Yes/No
4.	Technical Bid Compliance Proforma uploaded (Annexure-I).	Yes/No
5.	Whether an affidavit regarding Non-Black listing as per proforma given at Annexure-II duly attested by an Executive Magistrate or a Notary Public uploaded.	Yes/No
6.	In case the bidder is Authorized Supplier/Agency, the Authorization Certificate as per the Format given at Annexure-‘III’ uploaded.	Yes/No
7.	In case the Tenderer is Authorized Supplier/Agency, an undertaking/certificate issued by their Principle Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply then the Principle Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC (Annexure – ‘IV’) uploaded.	Yes/No
8.	Details of Bank Account for refund of EMD (Annexure – V) uploaded.	Yes/No
9.	Price Bid in the prescribed format in Excel Sheet (Annex – VI uploaded.	Yes/No
10.	Copy of Certificate of Registration for service Tax/TIN/TAN/PAN uploaded.	Yes/No
11.	A certificate from C.A. regarding Annual Turnover with Balance Sheet for the last 3 (three) financial years i.e. 2017-18, 2018-2019 and 2019-2020 uploaded.	Yes/No
12.	Copy of the IT Returns for three financial i.e. 2017-18, 2018-2019 and 2019-2020 uploaded	Yes/No
13.	Certificate regarding standard in quality .	Yes/No
14.	Guarantee , Projected delivery and Installation time, Make & Model	Yes/No
15.	Fax Number	
16.	E-mail ID	

Signature & seal of bidder

Place:

Date :

Note: Please upload Catalogue/Brochure/Pamphlets with complete specifications of quoted model only.

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.***

Annexure-II

**(To be furnished on non-judicial stamp paper worth Rs.100/- duly attested by
Executive Magistrate or Notary Public).**

AFFIDAVIT

I/We _____
partner/sole proprietor (Strike out which is not applicable) of (Name & Address of Firm)
_____ do hereby declare and solemnly affirm:-

- a) That the individual/firm/ companies are **not debarred or black- listed** by any department of Union/ State Government or any autonomous institute.
- b) That no partner or shareholder, directly or indirectly connected with the applicant who has been debarred or blacklisted by any department of Union Govt./State Govt./Autonomous Institute.
- c) And that the terms and conditions for supply and Installation of Equipments at GGSMCH, Faridkot, are acceptable to me/us. I/We shall abide by them in letter and spirit.

Date:

Place:

DEPONENT

VERIFICATION

I/We do hereby solemnly declare and affirm that the above declarations are true and correct to the best of my/our knowledge and beliefs. No part of it is false and nothing has been concealed therein.

Date:

Place:

DEPONENT

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.***

Annexure- III

MANUFACTURER'S/PRINCIPAL AUTHORIZATION FORM

TO

The Registrar
Baba Farid University of Health Sciences,
Faridkot -151203

Ref. No.....

Dated:

**Sub: Authorization Certificate in favour of M/s..... for supply
of (Name of equipment)**

We, M/s....., who are established and reputable
manufacturers of(name of equipment) having factory(ies) at
..... and, hereby authorize
M/s.....(name and address) to bid, negotiate and conclude
the Tender formalities with you against Tender No..... for the above
equipment(s) manufactured by us.

No company or firm or individual other than M/s..... are authorized
to bid, negotiate and conclude the tender formalities in regard to this business against this
specific tender.

We, hereby extend our full guarantee and warranty as per the conditions of tender for
the goods offered for supply against this tender by the above firm.

Yours faithfully,

(Name)

For and on behalf of M/s _____
(name of manufacturer/Principal)

**Note: This letter should be signed by a person competent and having authority to sign
on behalf of manufacturer, and should be on manufacturer Letter Head and same will
be uploaded with Technical Bid.**

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.***

Annexure - IV

UNDERTAKING BY MANUFACTURER/PRINCIPAL SUPPLIER

TO

The Registrar
Baba Farid University of Health Sciences,
Faridkot -151203

Ref. No.....

Dated:

Sub: Undertaking for after sales service

We, M/s....., who are established and reputable manufacturers of(name of equipment) have authorized M/s.....(name and address) to bid, negotiate and conclude the Tender formalities with you against Tender No..... for the above equipment(s).

Further, we undertake that in case dealership/distributorship is withdrawn after supply of equipment then we shall be responsible for after sales service till the date of guarantee/warranty of the equipment and afterwards for a period of 10 years.

Yours faithfully,

(Name)

For and on behalf of M/s _____
(name of manufacturer/Principal)

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.***

Annexure- V

Details of Bank Account of the firm who has deposited EMD

Name of the firm: _____

Sr. No.	Particulars	Detail
1.	Account No.	
2.	Name of Bank	
3.	Branch Name	
4.	IFSC Code of Bank	
5.	Name of Operator	

Baba Farid University of Health Sciences, Faridkot

*E-TENDER NOTICE FOR procurement and installation of Portable HF 100mA X-Ray Machine
required at GGS Medical College & Hospital, Faridkot.*

ANNEXURE - VI

PRICE BID

**TO BE UPLOADED in Printed/Computerized format in Excel Sheet Attached for all
Equipments.**