

GURU GOBIND SINGH MEDICAL COLLEGE, FARIDKOT. (PUNJAB)- 151203. (Constituent Medical College of Baba Farid University of Health Sciences, Faridkot)

E-mail: pr ggsmc@yahoo.com Website: www.ggsmch.org
Phone: 01639-251111 Fax: 01639-251070

No. Purchase/GGS/2020/		A :	Date:	19/3	190	9
	6 5 7			0	N 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	

Sub: Quotation for Supply of Light Source required at this Institution.

Sealed quotations are invited for supply of Light Source on terms & conditions given as under.

Sr. No.	Name of the Item	Specifications	Qty Required
. *	i a	**** *** *** *** *** *** *** *** *** *	
1.	Light Source with	Technical Parameter	S 15
	Fibreoptic Cable		As per
	Thoreoptic casis	External Power Supply: 100-240V/AC 50/60 Hz	Requirement
		Led power :120w	
		Color Temperature: 5000K-6500K	8 9
	Tree.	Intensity:100000 to 110400 lux	*
		Luminous Flux 2200 to 2484 Lm.	
		Noise: Less than 55 dB	
	28.58.85	Shell temperature: Less than 55° C	
	2 0 0 0	Safe type: I kind; BF type	
	A.	Working environment	x 2 - x - 7
		Temperature: 5~32° C	
	-	Relative Humidity : ≤70%	
		Atmospheric pressure: 86.0~106.0k Pa	
	41,	High Brightness is comparable to 350 W	
		traditional xenon light source, could be used for	
	A.S. Salar	laparoscopy without any proble,	
		Long life, it could last minimum.60, 000 hours.	
	n And NO	Power range: Wide range power from 100-240	
		V/AC; 60Hz.	and the second second
		Temperature:	9, 82
		Perfect color temperature 6000K to 6500 K,	
	v _a a	color rending is over 70K.	
	-	Does not emit light in UV or IR.	
		boos net that -g	2 2 2
		Customized connector, it accepts fiber light	
	8 2	guides with active areas ranging from 3 mm	
		to 10 mm diameter.	
	2. 240	to to min diameter.	·
		Energy saver for more than 50%	
	W	Environment friendly.	- s ⁻⁸
W	* = 0	Compatible with Karl storz, Olympus & stryker	
	a ·	camera system.	
	8	camera system.	. 8