

GURU GOBIND SINGH MEDICAL COLLEGE, FARIDKOT. (PUNJAB)- 151203. (Constituent Medical College of Baba Farid University of Health Sciences, Faridkot)

E-mail: procurement@ggsmch.org Website: www.ggsmch.org

Phone: 01639-251111 Fax: 01639-251070

No. Purchase/GGS/2021//2485	Date: 3//6/2021	1
	10//	

Sub: Quotation for Supply of Digital Colorimeter and pH Meter of wide range required at Biochemistry department.

Sealed quotations are invited for supply of Digital pH meter of wide range on terms &

Sr. No.	Name of the Item	Quantity Required
1.	Digital Colorimeter	As per
	Specifications:	requirement
	 Measuring Modes:%T & Abs., Conc. And K-factor 	
	Sample volume: 1ml Display:LCD	_
	◆ Light source:White LED	
	 Eight Optical filters on a turret to cover 400-700nm 	
	 The Wavelength Range should be 400-700nm with 8 optical filters 	
	with Peak wavelength 420nm, 440nm, 490nm, 520nm(typical)540nm,	
	600nm,700nm	
	Source:White LED	
	 Photometric Resolutionn:0.1% T & 0.001 Abs upto 	8
	1.999Abs(O.D.),Display:2Line 16 Character LCD	20
	 K-Factor Range: 1 to 1999, Concentration:0 to 2 	
	Photodiode detector	
	• Power 230V±10%,50Hz	-
	 Accessories : Set of 5 compatible tubes and Cuvettes 	
2.	Digital pH meter of wide range	
	Specifications:	
	PH:Range should be 0 to 14 with Resolution:0.1,0.01,0.001 Relative	2
	Accuracy:±0.002	
1	 Should have Up to 5 Calibration points with Auto & Manual 	
	calibration	
	 mV/RmV:Range should be 0 to ±1999.0mV with Resolution:0.1 mV 	
	& Relative Accuracy:±0.05% of reading	
	• Operatin Temp: Range: should be 0 to 99.9 C Resolution:0.1 C, 0.1 F	
	Relative Accuracy:±0.1	
	Input: pH Electrode:BNC, reference pin ATC Probe:8-pin min-DIN	
	Output:RS232	

The bidders should have to follow all the Terms & Conditions.

The Material should be of Good Quality and as per requirement.

2. The material should be strictly as per specifications. Certificate must be attached.

3. Supply should be F.O.R Destination i.e GGS, Medical Hospital, Faridkot.

- 4.Rates quoted should not be more than those quoted to **DGS&D** and any other Central or State Govts. Organizations.
- 5. Taxes should be mentioned extra as applicable.
- 6. The quantity may increase of decrease according to the requirement.

7. PAYMENT TERMS

The 80% payment will be made after successful installation/commissioning of machinery at the destination and remaining 20% payment will be released after 60 days of successfully running of the machinery/instrument and satisfactory performance. The material will be accepted only if it is found as per approved specifications.

- 8. **DELIVERY PERIOD**: Delivery & Installation Period:-within 30 days after issue of supply order.
- 9. **PENALTY CLAUSE**:If the supply is not made within the stipulated period then late delivery charges @2% will be imposed on the total amount of Supply Order up to delay of 30 days and thereafter @ 4% for another 30 days after which Supply Order will be deemed cancelled & security/earnest money forfeited and company will be black-listed for future.

10. ELIGIBILITY

- i) The sole manufacturers of equipments or their authorized agents/distributors may quote their rates.
- ii) In case of Authorized Supplier/Agency/Distributor, the Authorization Certificate should be attached.
- 11. The Quotation Biding Company will make arrangements for demonstration of equipment, if need be, at their own expenses.
- 12. The Quotation should certify unequivocally that the articles or items offered conform strictly to the specifications given by him. Any accessories, as may be required, should form part of the machine
- 13. Price of AMC/CMC should be quoted & it will be optional for institute to opt for any.

Note: Only Terms & Conditions mentioned on this Quotation will be considered for supply order.

(